



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/684,086	10/10/2003	Kimbolt Young	03-255 US	3032
23410	7590	11/02/2007		
Vista IP Law Group LLP 2040 MAIN STREET, 9TH FLOOR IRVINE, CA 92614			EXAMINER COHEN, LEE S	
			ART UNIT	PAPER NUMBER
			3739	
			MAIL DATE	DELIVERY MODE
			11/02/2007	PAPER

Please find below and/or attached an Office communication concerning this application or proceeding.

The time period for reply, if any, is set in the attached communication.

Office Action Summary

Application No.

10/684,086

Applicant(s)

YOUNG ET AL.

Examiner

Lee S. Cohen

Art Unit

3739

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☒ Responsive to communication(s) filed on 16 October 2007.
- 2a) ☒ This action is **FINAL**. 2b) ☐ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) 1-47 is/are pending in the application.
- 4a) Of the above claim(s) 6,8,18,20-22,24,31,33,40 and 47 is/are withdrawn from consideration.
- 5) ☐ Claim(s) _____ is/are allowed.
- 6) ☒ Claim(s) 1-5, 7, 9-17, 23, 25-30, 32, 34-39, and 41-46 is/are rejected.
- 7) ☐ Claim(s) _____ is/are objected to.
- 8) ☐ Claim(s) _____ are subject to restriction and/or election requirement.

Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on _____ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
a) ☐ All b) ☐ Some * c) ☐ None of:
- ☐ Certified copies of the priority documents have been received.
 - ☐ Certified copies of the priority documents have been received in Application No. _____.
 - ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).
- * See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- ☐ Notice of References Cited (PTO-892)
- ☐ Notice of Draftsperson's Patent Drawing Review (PTO-948)
- ☐ Information Disclosure Statement(s) (PTO/SB/08)
Paper No(s)/Mail Date _____
- ☐ Interview Summary (PTO-413)
Paper No(s)/Mail Date. _____
- ☐ Notice of Informal Patent Application
- ☐ Other: _____

Art Unit: 3739

DETAILED ACTION

Claim Rejections - 35 USC § 102

The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless –

(b) the invention was patented or described in a printed publication in this or a foreign country or in public use or on sale in this country, more than one year prior to the date of application for patent in the United States.

Claims 1-5, 7, 9-17, and 23 are rejected under 35 U.S.C. 102(b) as being clearly anticipated by Stadlmayr (DE 2,124,684). Applicant's attention is directed to the Figure 3 embodiment. The various electrodes are configurable to assume different active and common configurations; however, the term "configurable" relates to no more intended use of the electrodes absent any positively recited structure effecting the same.

Claim Rejections - 35 USC § 103

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

Claims 1-5, 7, 9-17, 23, 25-30, 32, 34-39, and 41-46 are rejected under either 35 U.S.C. 102(b) or 35 U.S.C. 103(a) as being anticipated by, or in the alternative, unpatentable over Mahvi et al (2002/0022864). The basic device is disclosed by Mahvi et al in Figures 2, 3, and 7. The use of a third electrode set is clearly shown in Figure 7 and detailed at paragraph [0069]. Figure 2 also shows the use of one of the electrode sets as a common electrode, while Figure 4 discloses that a large lesion volume is desired. The various electrodes are configurable

Art Unit: 3739

to assume different active and common configurations; however, the term “configurable” relates to no more than intended use of the electrodes absent any positively recited structure effecting the same. However, it would have been obvious to the skilled artisan to render the middle electrode as the common electrode since the reference clearly teaches larger lesion volumes are desired and such connection would inherently accomplish the same.

Response to Arguments

Applicant's arguments with respect to the above claims have been considered but are not deemed to be persuasive.

With respect to Stadlmayr, the Figure 3 embodiment is detailed to show the separate electrode elements are connected with opposite polarity. The reference also discloses that computer or other switching of the electrode elements is employed to obtain the desired switching. The electrodes are clearly configurable to effect such connection.

With respect to Mahvi et al, applicant's attention is directed to claims 17 and 20 in addition to the description of Figure 7 as detailed supra. Mahvi intends to use the third or common electrode to control the flow of current from the first and second electrodes. The electrodes are clearly configurable to effect such connection; however, it would have been obvious to the skilled artisan to use the middle electrode as the common electrode since a predictable result would ensue. The controller would be designed to effect the connection.

Allowable Subject Matter

Claim 19 is objected to as being dependent upon a rejected base claim, but would be allowable if rewritten in independent form including all of the limitations of the base claim and any intervening claims.

Art Unit: 3739

Conclusion

THIS ACTION IS MADE FINAL. Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).


A shortened statutory period for reply to this final action is set to expire **THREE MONTHS** from the mailing date of this action. In the event a first reply is filed within **TWO MONTHS** of the mailing date of this final action and the advisory action is not mailed until after the end of the **THREE-MONTH** shortened statutory period, then the shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of the advisory action. In no event, however, will the statutory period for reply expire later than **SIX MONTHS** from the mailing date of this final action.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Lee S. Cohen whose telephone number is 571-272-4763. The examiner can normally be reached on Monday-Friday, 7:00-3:30.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Linda Dvorak can be reached on 571-272-4764. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Art Unit: 3739

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.


Lee S. Cohen
Primary Examiner
Art Unit 3739

LSC
October 25, 2007